

	INSTRUCTION FOR USE	REF. NO.: RP.24.05-02 DATE: 20.04.2018
	PRODUCT : CAGES	REV. NO: 07 REV. DATE: 28.04.2025

This IFU is valid for all our brands: PRODORTH, PRD+, S33+
Important Information for the Surgeon!

➤ **OBJECTIVE :**

This device is used in the treatment of the anatomical abnormalities of vertebrae typically due to degenerative intervertebral discs. The device can be of several different geometric forms and is implanted between the vertebrae and providing mechanical stability and sufficient space for therapeutic spinal bone fusion to occur. This process helps to relieve pressure on pinched nerves and prevents vertebral slipping. For optimal results, a detailed preoperative evaluation, a meticulous surgical technique and adequate post-operative care are mandatory. It is important that both the patient and surgeon be fully aware of the risks and possible complications associated with this type of surgery. Before attempting this technique, the surgeon is advised to attend a training course with a surgeon already experienced with the use of the device.

➤ **DESCRIPTION :**

Prodorth cages are inserted into the disc space. The cervical cages are introduced by the anterior approach and lumbar cages are introduced by the posterior approach using special instruments. One or more cages may be required per segment for fusion in order to stabilize the segment concerned. Fusion is made between both vertebral endplates using bone grafts and the actual bone of the patient previously introduced into the cages. It is essential to insert implants with instrumentation specifically designed for this purpose. Prodorth cages are manufactured of Titanium alloys as well as the PEEK material due to their high bio-compatibility and to the fact that they may be imaged better in such radiological examinations as MR, CT, etc. FOR USE BY A PHYSICIAN, SURGEON OR SPECIALIST DOCTOR ONLY.

➤ **IMPLANT MATERIAL:**

The raw materials used in the Prodorth Cervical Disc Prosthesis are VESTAKEEP PEEK by EVONIK INDUSTRIES as indicated by the symbol “” (ASTM F-2026) also the titanium alloys (ASTM F-136). Prodorth expressly warrants that these devices are fabricated from a combination of the foregoing material specifications.

➤ **INDICATIONS:**

Lumbar cages,
• Designed for use with a posterior approach at or near levels from L2 to S1 for the treatment of (DDD)
• Degenerative disc pathologies
• Herniated nucleus pulposus
• Grade 1 degenerative or isthmic spondylosis
• Visible loss of disc height compared to adjacent levels
• Lumbar pseudarthrosis
Cervical cages,
• Designed for use from C3 to C7 with an anterior approach, for the treatment of cervical degenerative disc disorders
• In the treatment of lack of stability in the disc and vertebra
• Cervical pseudarthrosis
Note: Patients should be skeletally mature and have had six months of non-operative treatment.
• It is recommended to place bone graft into the implant

➤ **CONTRAINDICATIONS :**

Prodorth Cages should never be used at any condition not described in the indications for use. Contraindications include, but are not limited to:

PRODUCT : CAGES

- Fracture, tumor
- Any abnormality present which affects the normal process of bone remodeling including, but not limited to, severe osteoporosis involving the spine, bone absorption, osteopenia
- Marked local inflammation
- Pregnancy
- Infection

- Recognized allergies to titanium or titanium alloys and PEEK material
 - Damaged vertebrae from an accident (trauma) at the level of the surgery
 - Prior fusion at the level(s) to be treated
 - An unhealthy shape (deformity) of the vertebrae at the level of the surgery
 - Low bone mineral density, such as osteoporosis or osteopenia
 - Mental disability
 - Obesity
 - Open wounds
 - Fever or leukocytosis
 - Alcohol or drug addiction
 - Uncooperative patient or patient with neurologic disorders rendering the patient incapable of following instructions
- These contraindications can be relative or absolute and should be considered when physician makes a decision. The above list does not include all possibilities. Surgeons should discuss relative contraindications with the patient.

- An unhealthy shape (deformity) of the cervical vertebrae at the level of the surgery
 - Low bone mineral density, such as osteoporosis or osteopenia
 - Severe facet joint disease or degeneration
 - Mental tdisability
 - Any condition not described in the indications for use
 - Obesity
 - Open wounds
 - Fever or leukocytosis
 - Alcohol or drug addiction
 - Uncooperative patient or patient with neurologic disorders rendering the patient incapable of following instructions
- These contraindications can be relative or absolute and should be considered when physician makes a decision. The above list does not include all possibilities. Surgeons should discuss relative contraindications with the patient.

➤ **SECONDARY AND POSSIBLE SIDE EFFECTS :**

The patient shall be notified regarding the below mentioned adverse events pre-operatively. A second surgical treatment may be required:

- Pseudarthrosis
- Implant penetration, migration or Implant failure
- Infection
- Tissue or nerve damage caused by improper positioning and placement of implants or instruments
- Paralysis
- Allergy to materials used
- Dysphagia (related with cervical levels)
- Loosening
- Increased pain
- Instability
- Hematoma
- c7 palsy (related with cervical levels)
- Hoarseness (related with cervical levels)
- Displacement of the disc adjacent segment degeneration
- Pain or illness
- Nonunion or delayed union of the bone
- Bleeding blood vessels

PRODUCT : CAGES

- Inability to perform daily activities
- Dura leak requiring a repeating surgery
- Intervertebral cage can be fractured postoperatively above or below segments of the surgical level due to trauma, presence of any defect or weak bone structure. Re-operation may be required
- Wound infection
- Death



➤ **CAUTIONS OF USE :**

Never re-use an implant even in a perfect state. Any Implant which has been used, twisted, bent, implanted and then removed even If it appears intact must be discarded.
Use new implants routinely.

Correct selection of the implant is highly important! Use of provided trials is recommended.

➤ **RESIDUAL RISKS**

- In case of working with a supplier that does not meet raw material standards, the raw material to be purchased may not meet both the performance requirements of the product and the safety requirements.
- Inappropriate product production due to inadequate measurement devices in the control section may cause inadequate performance and customer dissatisfaction.
- Failure to properly clean the machine oils that the product is exposed to during machining may lead to an unsuitable product, delaying the assembly process and delaying the shipment at the same time. This may lead to customer dissatisfaction.
- The shipment may be delayed due to assembly errors or an unsuitable product may be shipped. Customer dissatisfaction or incorrect implantation may occur as a result of inappropriate product shipment.
- Biomechanical strength may decrease due to load imbalance as a result of implant misalignment during the operation and this may cause the patient to undergo a re-operation.
- Excessive load on the implant before the implant-superior/inferior vertebral pads are fully attached may result in bending or fracture of the implant due to excessive pressure created at the vertebra-implant-vertebral interface.
- Chemical contamination present on the product may cause complications in the patient due to toxic effects.
- Non-asepticity of the area/people who will perform the operation may expose the patient to the risk of deep infection. As a result of the infection in question; the patient may face a life-threatening situation.



➤ **WARNINGS :**

Preoperatively:

The surgeon must be fully conversant with all aspects of the surgical technique and know the indications and contra-indications of this type of implant. The surgeon must have acquainted himself before the operation with the specific technique for insertion of the product which is available from the manufacturer. As part of the pre-operative examination, the surgeon must check that no biological, biomechanical or other factors will affect the correct conduct of the operation and the post-operative period. An appropriate range of sizes must be available at the time of the operation. Patient must be warned beforehand about not forcing the implant before a complete fusion.

PRODUCT : CAGES

Prodorth cages can be broken when subjected to the increased loading associated with delayed union or non-union. Notches, scratches or bending of the implant during the course of surgery may also contribute to early failure. Hence a visual examination of the implants is highly important before the surgery. The correct selection of the type and the size of the implant, as well as the positioning are extremely important.

Preoperative planning may be used to estimate the required implant size and to assure that the appropriate range of sizes are available for surgery. All implants and instruments shall be opened, visually controlled for possible damages, cleaned and sterilized preoperatively. If there are some disorders about surface smoothness, do not use the implant and contact to supplier. A surgical technique for the Prodorth Cages is available upon request. This technique is not a substitute for training and is for general informational purposes only.

PHYSICIAN NOTE : Although the physician is the learned intermediary between the company and the patient, the important medical information provided in this document should be explained to the patient.

Postoperatively: Patients must be informed of the precautions to be taken in their everyday life to guarantee a maximum implant service life. It is recommended that a regular postoperative follow-up be undertaken to detect early signs of failure of the implants and to consider the action to be taken. Deterioration of the device after bone consolidation cannot be considered to constitute a dysfunction or deterioration in the characteristics of the implant. A suitable rehabilitation program must be designed and implemented.

➤ HANDLING AND STORAGE :

The handling and the storage of the cages can be at room conditions. The implants must be stored with care. Should these requirements not be followed, reduced mechanical properties may occur, which could lead to implant failure in some cases. Proper function of the surgical instruments specific to the cages should be checked prior to use.



➤ CLEANING - DECONTAMINATION :

Prodorth cages are not supplied sterile. An appropriate sterilization method should be used. All packaging and labeling must be removed before the next steps. The cleaning and decontamination must be completed before sterilization. All instruments and implants must be disassembled (if applicable) and then cleaned and disinfected using neutral cleaners before sterilization and introduction into a sterile surgical field. Cleaning and decontamination must include the use of neutral cleaners followed by a deionized water rinse. Cleaners and disinfectants should be intended to disinfect the medical device and CE certified as well. Never use metallic brushes for cleaning which may damage the products. Cleaning in a machine with products adapted and dry all products which can alter the implants are forbidden.

PRODUCT : CAGES

➤ STERILIZATION :

Prodorth Cages are released to market as non-sterile. Prodorth Cage as well as its instruments must be sterilized by hospital prior to surgical use. All packaging materials are removed prior to sterilization. The recommended sterilization method for PRODORTH cages is steam sterilization in an autoclave. The products which are intended to be sterilized should remain in an autoclave at 134 °C for 18 minutes. There is no other sterilization method PRODORTH recommends.

Note: Due to many variables in sterilization, each medical facility should calibrate and verify the sterilization process (e.g., temperatures, times) used for their equipment.

Use appropriate protective wrapping and add the original stickers preventing the implant from the direct contact with the sticker. Care should be taken to protect parts from mechanical damage. The recommendation given is for information only. The manufacturer and distributor assume no responsibility for Prodorth products for a improper sterilization by the user.

➤ ATTENTION :

Similar products of competitors shall not be combined with the components of the fusion cages. Prodorth implants and instruments should only be used with Prodorth instruments. In case of using other company's instruments, this might result in galvanic corrosion, incompatibility between the products as well. No component of the Prodorth Cages System implants shall be reused. The restricted shelf life of the device is 10 years. It should never be used after its expiration date.

➤ DISPOSAL :

Implants removed from the patient at revision surgery should never be reimplanted as the internal structure of the implant cannot be determined by visual examination. Removed implants must be treated as biological hazards and required to be treated or disposed of according to the country's waste regulations, hospital policies and procedures where the implant is removed.

➤ PACKAGING :

Implants and instruments are delivered inside the special trays of instrument boxes, non-sterile packages or in locked bags. All products shall be controlled and accepted by the receiver. If there is any damage on the outer packaging please return the relating product to RD MEDİKAL INC. immediately.

➤ TRACEABILITY :

There is always a lot number on the label of each package or over implant. This lot number must be attached to the le of the patient in order to trace back for production procedures. Because of traceability reason, distributional documents have to be kept for 15 years.

➤ PRODUCT COMPLAINTS :

Any health professional (e.g. surgeon using the products) who has a complaint or is dissatisfied with quality, identification, reliability, safety, efficacy and/or performance of Prodorth Cages must inform either Prodorth or the distributor. If there is a serious adverse event or risk of such, liable to cause death or having caused death or serious problem in the state of a patient or patient's health, Prodorth (or the distributor) must be informed immediately by phone or mail. All complaints must be accompanied by the product name, ref number and lot number of the component. The person formulating the complaint should state the name, address and the nature of complaint, giving as many details as possible.

➤ FURTHER INFORMATION :

In case of complaint, or for supplementary information, or further directions for use of this system, please see the address listed below.

PICTOGRAM	EXPLANATION
	DO NOT REUSE
	NON-STERILE
	NOTIFIED BODY
	DO NOT USE IF DAMAGED
	CATALOGUE NUMBER
	PRODUCTION DATE
	EXPIRE DATE
	CONSULT INSTRUCTION FOR USE
	LOT NUMBER
	CAUTION
	MANUFACTURER
	KEEP AWAY FROM SUNLIGHT
	COUNTRY OF MANUFACTURE
	KEEP AWAY FROM RAIN
	MEDICAL DEVICE
	MR CONDITIONAL

Please check our website for the latest version of this IFU.



RD Medikal Tıbbi Ürünler san. ve Tic. A.Ş

Karacaoğlan Mah. Bornova Cad. No:9G/1 Bornova İzmir / TURKEY

T: 0090.232.348 4950 (Pbx)

info@prodorth.com